



Spojené ovládání
Elektrické připojení ca. 5 kW.
400V / 25A / 50Hz / 3-N-PE.
Elektrické jistiění (automatickým jističem).
Propojení do strojí provede Jensen Seaking

Spínač panel -na každém sístiči-
Elektrické připojení = 11,3 kW
400V / 50A / 50Hz / 3-N-PE
Externí jistiění (automatickým jističem)

- | Poz. | Ks. | Popis |
|-------|-----|--|
| 1 | 1 | Nakládací dopravník, 4 komorový x 36 kg, stávací (2013885) |
| 2 | 1 | Kontinuační pračka P36-13 Uni, stávací (2013885) |
| 2a | 1 | Programovací panel s monitorem, stávací (2013885) |
| 2b | 1 | Přídavný tank na vodu, stávací (2013885) |
| 3 | 1 | Dávkovácí lis SEP 36 MD, stávací (2013885) |
| 4 | 1 | Rozváděcí dopravník 1000 x 1800 mm AA, stávací (2013885) |
| 5.1-3 | 3 | Průchozí sušič DT 40 U, stávací (2013885) |
| 6 | 1 | Výška podstavce 1780 mm |
| 7 | 1 | Spojené ovládání, stávací (2013885) |
| 8 | 1 | Vykádací dopravník 800 x 6500 mm AA, stávací (2013885) |
| 9 | 1 | musí být modifikováno, nová délka 8500 mm |
| 11 | 1 | Třídící dopravník 800 x 4000 mm AA, stávací (2013885) |
| 20 | 1 | Dřevěné hrzení,
musí být modifikováno místně.
Průchozí sušič DT 60,
výška podstavce 1780mm
Sasí systém pro Poz. 11 |

The description shown requires a load headroom of 5000 mm.
Dimensions for rigging:
width = 2200 mm, height = 3000 mm.

The area under the base frame has to be shut by means of a safety device by other when a continuous batch washer has a pedestral height of more than 180 mm.
The given dimensions for existing machines, devices, supply lines and hoses are without liability. The receiver is asked to once again check these data!
Execution of the installation parts see in the offer and order texts respectively.
Protective devices like grids etc. and working platforms have to be provided locally according to the local safety regulations for prevention of accidents.
Warning! Do not scale this drawing. Verify all dimensions and structural details on site. If in doubt, ask before scaling.

B	10-10-25	AE	New position suction nozzle
A	10-10-21	AE	New position suction nozzle, modified extended conveyor
REV :	DATE :	CD :	Status/Amendments :
COPYRIGHT			
The copyright administrator and user of this document shall be the manufacturer of its content as the manufacturer of the equipment. The copyright shall be granted to the manufacturer of the equipment in the event of the grant of a patent, utility model or design.			
Fakultní nemocnice Olomouc		ALL DIMENSIONS IN mm	
Olomouc, CZ		Checked By:	
SSC / Factory:		Date Checked:	
Date: 2010-10-12		Name: AE	
Order No: 2072303/001		Copyright recording to be observed!	
REVISION: B		Date:	
SCALE: 1 : 50		Dwg. No: E - 1748 B	