



MANUFACTURERS EC DECLARATION OF CONFORMITY

Kortrijk, 29 September 2017

Manufacturer Identification:

Legal Name : Barco NV
Address : President Kennedypark 35
8500 Kortrijk
Belgium

Representative : Patrick Santy
Function : Vice President NPI Excellence

Product Identification:

Product : LCD Monitor
Brand : BARCO
Model : **MDRC-2222**
Type number : MDRC-2222 BL; MDRC-2222 TS; MDRC-2222 WP; MDRC-2222 WH

CE affixing date : 2017

Declaration:

Provided that it is installed, maintained and used in the application for which it is made, with respect of the professional practices, relevant installation codes and manufacturer's instructions:

*We hereby declare, under our sole responsibility, that the above referenced product complies with the essential requirements of Council Directives
2014/30/EU (EMC), 2011/65/EU (RoHS) and 93/42/EEC, amended by 2007/47/EC class I (Medical Devices)*

The procedures of Annex VII have been applied to mark the product with the CE-label.

European harmonised standards applied:

EMC: EN 60601-1-2 (2015)

Product safety: EN 60601-1 (2006) + A1 (2013) + A12 (2014)

RoHS: EN 50581 (2012)

Signature of manufacturer representative:

Patrick Santy





(prostý překlad)
ES Prohlášení o shodě

Kortrijk, 29. září 2017

Identifikace výrobce:

Název: Barco NV
Adresa: President Kennedypark 35
8500 Kortrijk
Belgie

Zastupující: Patrick Santy
Funkce: více prezident NPI Excellence

Identifikace výrobku:

Výrobek: LCD Monitor
Značka: BARCO
Model: MDRC-2222
Typové číslo: MDRC-2222 BL; MDRC-2222 TS; MDRC-2222 WP; MDRC-2222 WH

Datum označení CE: 2017

Prohlášení:

Za předpokladu, že je nainstalován, udržován a používán v aplikaci, pro kterou je vyroben, je dodržována odborná praxe, příslušné instalační kódy a pokyny výrobce:

Tímto prohlašujeme, na naši výhradní odpovědnost, že výše uvedený výrobek vyhovuje základním požadavkům směrnice **Rady 2014/30/EU (EMC), 2011/65/EU (RoHS) a 93/42/EHS, ve znění směrnice 2007/47 /ES třídy I (Lékařské přístroje)**

Pro označení produktu štítkem CE byly použity postupy uvedené v příloze VII.

Použité evropské harmonizované standardy:

EMC: EN 60601-1-2 (2015)

Bezpečnost výrobku: EN 60601-1 (2006) + A1 (2013) + A12 (2014)

RoHS: EN 50581 (2012)

Podpis zástupce:

(nečitelný podpis)
Patrick Santy

CE prohlášení Zdravotnické prostředky EN 60601 třída I – vydání 13