



**Test Protocol**  
**Electrical Safety Test of Endoscopes**

<b>Repair Number:</b> 2007/1689		<b>Location:</b> HH	
<b>Customer Number:</b> 00098892 <b>Customer:</b> FN Olomouc		<b>Service Organisation:</b> Olympus Brno National	
<b>Serial Number:</b> 2501668		<b>Model:</b> GIF-Q180	
<b>Testing Person:</b> Karny		<b>Date of Test:</b> 9.4.2008	
<b>Income Inspection:</b> [   ]		<b>Outgoing Inspection:</b> [ X ]	
<b>Measurement Function</b>	<b>Meas. Value</b>	<b>Limit Value</b>	<b>Result</b>
Pressure	2,4 mbar	< 3 mbar	PASSED
C <sub>X</sub> Water	2,6 nF	< 6,555 nF	PASSED
R <sub>INS</sub> Water	9,8 MOhm	> 1 MOhm	PASSED
Z <sub>SCOPE</sub>	11,9 Ohm	< 45 Ohm	PASSED
<b>Result</b>			<b>PASSED</b>
<b>Remark to the test:</b> For HF applications ensure an adequate and reliable connection according to manufacturer+s advice. The HF output- monitoring is mandatory and the output power shall not exceed the recommended limit for flexible endoscope applications.			
	<b>MPE Tester</b>	<b>LSM Tester</b>	
<b>Manufacturer</b>	GOSSEN-METRAWATT	Gigahertz Optik	
<b>Serial Number</b>	1234	0	
<b>ID Number</b>	12345		
<b>Firmware Version</b>	Endotest MPE V1.09		
<b>PC-Software Version:</b>	2.0.53		
<b>- This is an automated document , and therefore valid without being signed -</b>			
Location: HH		Date: 09.04.2008	

Copyright: