

Fakultní nemocnice Olomouc  
I.P.Pavlova 6  
775 20 Olomouc  
IČO: 00098892

### SERVISNÍ LIST

Dne 29.5.2007 byly na TK holteru Cardiette bp one sériové číslo AENG 0021 provedeny následující výkony:

- kalibrace základní desky
- kompletní prohlídka přístroje
- test těsnosti
- výměna konektoru
- výměna pumpy

Podrobnosti o provedených měřeních a kompletní výstup z kalibračního přístroje jsou součástí originálu "Detail list of executed performances".

S pozdravem,  
Irena Sedmíková  
MEDATRON, spol. s r.o.  
Sirotkova 45, 616 00 Brno  
Telefon: +420 541 235 672  
Fax: +420 541 246 681/9  
irena.sedmikova@medatron.cz  
www.medatron.cz



MEDATRON, spol. s r.o.  
Sirotkova 45, 616 00 Brno  
tel.: 541 240 838

FNOL0012006

et medical devices S.p.A

DETAILED LIST OF EXECUTED PERFORMANCES

SPE

N. ID 148	Repair. N°: 10700	Customer 100391023	MEDATRON SPOL S R.O.		
Q.ty 1	67010303	S/N AENG0021	REGISTRATORE BPONE		
Repair material documentation			Date acceptance of the goods	Repaired in warranty	
Blading bill	24052007	Date of B.B.	24-05-07	29-05-07	No

<b>DENOUNCED FAILURE:</b> GENERAL CHECK	<b>NOTICED FAILURE:</b>
--	-------------------------

**PERFORMED REPAIR:**  
Replacement of valve, calibration of mother board, general check and leakage current test.

**EXECUTED PERFORMANCES**

Troubles shooting	<input type="radio"/> No <input checked="" type="radio"/> Si	Test	<input type="radio"/> No <input checked="" type="radio"/> Si	Running in	<input type="radio"/> No <input checked="" type="radio"/> Si
Modification performed	<input checked="" type="radio"/> No <input type="radio"/> Si	General control	<input type="radio"/> No <input checked="" type="radio"/> Si	Safety tests	<input type="radio"/> No <input checked="" type="radio"/> Si

**REPLACED SPARE PARTS**

Part number	Description	Q.ty

LABOR	P/N 66700001 :	EQUIVALENT TOTAL QUANTITY	
MATERIAL	P/N 66700002 :	EQUIVALENT TOTAL QUANTITY	

**REPAIR WARRANTY ( 3 months ) :**  
THE REPAIR ON COMPLETE APPARATUSSES (FINISHED PRODUCTS) NOT UNDER WARRANTY, WILL BE COVERED BY A SUPPLEMENTARY WARRANTY OF 3 (THREE) MONTHS EXCLUSIVELY ON THE PART THAT HAS BEEN REPAIRED, WHICH WILL BE DATED, VALIDATED AND SIGNED BY THE AFTER SALES SERVICE MANAGER.  
FOR MORE SPECIFIC WARRANTY CONDITIONS, REFER TO THE WARRANTY CONDITIONS THAT ARE SUPPLIED TOGETHER WITH THE PRODUCT.

**NOTES :**

Service operator : *G. Gatti* Date : 15-06-07

PROTOCOLLO MISURAZIONI TEST FINALE REGISTRATORE BPONE VALVOLA 33 OHM

Data : 13. 6.2007

Collaudatore : PG

Numero seriale : 0001137

POS	MISURA BASSA	VALORE MINIMO	MISURA OK	VALORE MASSIMO	MISURA ALTA	UNITA'
1		Cambio di modo o.k.				
2		100.000	139.000	400.000		1
3		0.100	1.547	6.000		mmHg/min
4.1		10.000	17.110	20.000		sec
4.2		3.500	4.082	6.000		mmHg/sec
4.3		0.000	43.000	1500.000		1
4.4		15000.000	20600.000	32000.000		1
4.5		17600.000	20600.000	23600.000		1
5.1		79.000	80.000	81.000		l/min
5.2		118.000	122.000	126.000		mmHg
5.3		90.000	92.000	97.000		mmHg
5.4		72.000	80.000	83.000		mmHg
5.5		4000.000	4393.000	6100.000		1
6	Numero seriale della scheda madre = 451143					
8	Lotto di produzione della valvola = 0703					
9	Versione del software applicativo 02.12					
10	Nuovo numero seriale 0001137 programmato					
11	Parametri di reset o.k.					

\*\*\*\*\* Registratore collaudato conforme \*\*\*\*\*