

CE 0123

EC DECLARATION OF CONFORMITY



Manufacturer's Name: Medtronic Emergency Response Systems, Inc.

Manufacturer's Address: 11811 Willows Road NE
Redmond, WA 98052-2003 USA

OPIS

declares that the CE-marked product

Product Name: LIFEPAK® 20 defibrillator/monitor

Part Number(s): 3200500

complies with 93/42/EEC (Medical Device Directive) class IIb. Conformity assessed per Annex II.

This product complies with:

Safety: EN60601-1-1:1996/IEC 60601-1:1995
CLASS I, type BF with CF parts/Continuous operation
IEC 60601-2-4:1983
EN 60601-2-25/IEC 60601-2-25:1993
UL 2601-1:10-24-97*

EMC: EN60601-1-2:1993
EN 55011:1991 Class B, Group 1
EN 61000-4-2/IEC 61000-4-2:1995 3kV CD, 8kV AD
EN 61000-4-3:1996/IEC 61000-4-3:1995 3 V/m
EN 61000-4-4/IEC 1000-4-4:1995 0.5 kV Power Lines
EN 61000-4-5/IEC 1000-4-5:1995 0.5 kV Power Lines
and per EN 60601-1-2

*Tested according to Figure 11 of UL 2601-1:10-24-97

Supplementary Information

Included are the following accessories and interconnecting cables:

QUIK-COMBO™ pacing/ defibrillation/ ECG electrodes	Internal handles with discharge controls
QUIK-COMBO PEDIATRIC pacing/ defibrillation/ECG electrodes	Serial cable (system connector)
QUIK-COMBO RTS pacing/ defibrillation/ECG electrodes	FAST-PATCH defibrillation cable
QUIK-COMBO pacing/defibrillation/ ECG electrodes with REDI-PAK™ preconnect system	3-lead ECG cable, 5-lead cable
FAST-PATCH® pacing/defibrillation/ ECG electrodes	QUIK-COMBO defibrillation cable
FAST-PATCH PLUS pacing/ defibrillation/ECG electrodes	SpO2 cable PC04 (4 feet), PC08 (8 feet), PC12 (12 feet)
Standard paddles with built in pediatric paddles (two required)	SpO2 sensor, hard shell, finger (adult and pediatric)
External sterilizable paddles	SpO2 Disposable sensors (Masimo® compatible) (adult and pediatric)
Posterior paddles	SpO2 Option
	External Pacing Option
	Docking Station
	QUIK-COMBO Test Plug


Michael D. Willingham
Vice President, Regulatory Affairs

Redmond, October 25, 2004

This declaration applies to CE marked devices produced after the date of issuance of this declaration and before it is either superseded by another declaration or withdrawn.

Authorized EC Representative: Medtronic B.V., Earl Bakkenstraat 10, 6422 PJ Heerlen, The Netherlands

PŘEDÁVACÍ PROTOKOL – ZÁRUČNÍ LIST

Zákazník:

Č. zakázky:

Č. kupní smlouvy/objednávky:

FN Olomouc
 KOG0043
 ARK - klinika

POŘ. Č.	NÁZEV	TYP	KS	VÝROBNÍ ČÍSLO	ZÁRUKA
1.	UFEPAK 20	TACE	1	339 47 895	24 měsíců
2.	PADLA		1		
3.	POSLSUŠEČNÍK		1		6 m B
4.					
5.					
6.					
7.					
8.					
9.					
10.					
11.					
12.					
13.					
14.					
15.					
16.					
17.					
18.					
19.					
20.					

Předávaná dokumentace:

manuál k obsluze v českém jazyce
 prohlášení o shodě

Podpisem tohoto dokumentu potvrzují obě strany souhlas s níže uvedenými závazky a záručními podmínkami:


- Předávající ručí za kvalitu, bezpečnost a funkci přístroje/přístrojů po dobu uvedenou v záhlaví tohoto dokumentu.
- Během záruční doby se předávající zaručuje poskytnout zdarma veškeré opravy včetně potřebných náhradních dílů. Záruka se však nevztahuje na vady způsobené neodbornou obsluhou, poškozením, použitím přístroje k jiným účelům, než byl určen, použitím neoriginálního spotřebního materiálu a příslušenství či náhradních dílů nebo neodborným servisem nebo nadměrným opotřebením.
- Po uplynutí záruční doby se předávající zavazuje poskytnout na výrobek odborný servis za úplatu po dobu nejméně 10 let.

Přístroje předány dne: PURO KLIMA a.s. (6)

20.6.2006

Štěpánovická 2266/2
 100 00 PRAHA 10
 IČ 00149931 BIC 2200149931

Předávající
 Jméno, podpis, razítko



Přebírající
 Jméno, podpis, razítko

Společnost je zapsána do kolektivního systému zpětného odběru elektrozařízení ASEKOL.

Objednatel si vyhrazuje právo proplácet fakturované částky subdodavatelů v souladu s podmínkami uzavřených SoD s investorem nebo vyšším dodavatelem, včetně jimi určených dob splatnosti. Na základě této objednávky prosíme o předložení návrhu kupní smlouvy nebo smlouvy o dílo.

Fakturu odešlete samostatně poštou na adresu PURO-KLIMA a.s. a přiložte kopii naší objednávky.