

Fakultní nemocnice Olomouc
I.P.Pavlova 6
775 20 Olomouc
dispečink oprav

Protokol: O kontrole zdravotnického prostředku dle Zákona č. 346/2003 Sb. § 52

Sallanches, 27.6.2007

Vážení,

my, níže podepsaná firma Anthogyr, jsme na základě požadavku našeho zástupce pro Českou Republiku, firmy JANDA-DENTAL s.r.o., provedli kontrolu implantologického motoru Minighyr Vario, výrobní číslo 93718.

Výsledkem této kontroly je závěr, že tento motor vyhovuje a je způsobilý k dalšímu užívání v implantologii. Motor byl testován z hlediska funčnosti a elektrické bezpečnosti.

U tohoto motoru, jehož stáří je 13 let, byly na základě Vašeho souhlasu vyměněny některé součástky. Jedná se o ložiska, kartáče, části konektoru pedálu a některé spoje.

Pro další servis doporučujeme příští kontrolu za dva roky, tj. v červnu 2009.

V případě jakýchkoliv dotazů nás neváhejte kontaktovat.

S pozdravem

Serge SERRA
Pozáruční servis a opravy Anthogyr


Luboš Janda
JANDA-DENTAL s.r.o.

JANDA-DENTAL s.r.o.
Osvoboditelů 1175, 440 01 Louny
IČ: 25497359, DIČ: CZ25497359
tel.: 415 653 201, www.janda-dental.cz

Příloha: originál Protokol o kontrole



anthogyr



**JANDA-DENTAL
Mr Lubos Janda
Osvoboditelu 1175
440 01 LOUNY
CZECH REPUBLIC**

Protokol : Medical Device test report according CZ Law Nr. 346/2003 Sb. § 52

Sallanches, July 12th 2007

Dear Sirs,

We , undersigned Anthogyr, declared having controled the motor Minighyr N° 93718 on behalf of the company JANDA DENTAL s.r.o., located in Louny, Czech Republic.

This evaluation has led us to the conclusion that the motor is conform to an usage in implantoloy. This motor was tested in respect of the utility and electric safety.

This motor, which is 13 year old, has been fixed with your approval. The parts that have been replaced are ball bearings, brushes, screw form the connector of the pedal,as well as some weldings.

For a continuous maintenance of this motor Minighyr, we recommend a next control in two years, i.e June 2009.

We stay at your disposal for any further information, do not hesitate to contact us,

Best regards ,

Serge SERRA
After Sales Service Manager

ANTHOGYR

164, rue des Trois Lacs - 74700 SALLANCHES - France
Tél. +33 (0)4 50 58 02 37 • Fax +33 (0)4 50 93 78 60
www.anthogyr.com • E-mail : sales@anthogyr.com

DENTAL CARE BY NATURE

