

DECLARATION OF CONFORMITY

according to ISO/IEC Guide 22 and EN 45014

Manufacturer's Name: Hewlett-Packard GmbH

Manufacturer's Address: Hewlett-Packard-Strasse 8
D-76337 Waldbronn
Germany

declares that the product

Product Name: 8453 UV/Visible Spectrophotometer

Model Number(s): G1103A

conforms to the following Product Specifications:

Safety: IEC 1010-1:1990 + A1 / EN 61010:1993
CSA C22.2 No. 1010.1-92
UL 3101-1

EMC: CISPR11:1990 / EN 55011:1991 - Group 1 Class B^[1]

EN 50082-1: 1992

IEC 801-2:1991 - +/- 4kV CD, +/- 8kV AD

IEC 801-3:1984 - 3 V/m 80% sine mod.

IEC 801-4:1988 - 1kV Power Lines / 0.5kV Signal Lines

Supplementary Information:

The product herewith complies with the requirements of the EMC Directive 89/336/EEC and the Low Voltage Directive 73/23/EEC and carries the CE marking accordingly.

^[1] The product was tested in a typical configuration with Hewlett-Packard test systems.

Waldbronn, Germany, December 1995


(Wolfgang Wilde, Quality Manager)

European Contact: Your local Hewlett-Packard Sales and Service Office or Hewlett-Packard GmbH, Department ZQ/Standards Europe, Herrenberger Strasse 130, D-71034 Boeblingen (Fax: +49-7031-14-3143)

P/N: G1103-90501



Revision: B.00.00
effective by: 11 Dec 95

Document-No.:
ISO-QA05-1103-AP